



3-5 GPM
@ 15 PSIG (MIN)

1' 3-WAY BALL VALVE, T-PORT
REQUIRES 60 PSIG AIR, 1/4" NPT

1" NPT

FLUSH WATER TO
BACKFLUSH SENSOR
(+15 PSIG OVER
FLUID PRESSURE)

CT03/25
SENSOR

CSP-UP
CONSISTENCY
SIGNAL
PROCESSOR

30' POWER TO SOLENOID
110/1/60

33' CABLE

4-20mA
SIGNAL OUTPUT

POWER
110/1/60

Cerlic Environmental Controls, Inc

P.O. Box 420097
Atlanta, GA 30342
Phone (404) 256-3097
Fax (404) 256-3094



CERLIC HIGH SOLIDS METER

CT03/25 W/ 3-WAY
FLUSH VALVE

DRAWING NUMBER 107
PROJECT NUMBER CT03/25
SCALE 1/4

NO.	DATE	DESCRIPTION	BY
1			
2	02/26/02		JR
3			
4			
5			
6			
7			